


X-ray Blood Irradiators



CE 0459



Raycell[®] Mk1 X-ray Blood Irradiator

- Self-contained cooling system, no external water required
- Canister volume 1.6 L
- Estimated number of bags per cycle — 3 bags, 300 ml each
- Time for 25 Gy central dose ~5.3 min
- Time for 25 Gy minimum dose ~6.3 min
- Dose uniformity ratio 1.55:1
- Barcode reader
- Syringe holder



CE

Raycell[®] Mk2 X-ray Blood Irradiators — 2.0 L or 4.8 L (6 Canister) Models

- Two opposing x-ray tubes deliver excellent dose uniformity
- Central dose rate (+/- 5%)
2.0 L model: 12.3 Gy/min
4.8 L model: 5.8 Gy/min
- Irradiation time is controlled by two independent microprocessors
- Fast irradiation times for 25 Gy central dose
2.0 L model: ~2.0 minutes
4.8 L model: ~5.0 minutes
- Barcode reader



CE 0459

Raycell[®] X40 Research Irradiator

- X-ray version of the Gammacell[®] 40 Exactor
- Delivers a central dose of ~1.5 Gy/min
- 8.0 L sample container
- Excellent dose uniformity
- Self-contained cooling system — no external water required

Best Theratronics Ltd. 413 March Road, Ottawa, Ontario K2K 0E4 Canada
tel: 613 591 2100 866 792 8598 www.theratronics.com www.teambest.com

AFRICA | ASIA | EUROPE | LATIN AMERICA | MIDDLE EAST | NORTH AMERICA

European Representative: Emergo Group, Prinsessegracht 20, 2514 AP, The Hague, Netherlands